

Certificate of Compliance



No. 1S190123.NM0Q10

Certificate's
Holder:

NUGA MEDICAL Co., Ltd.
185, Jiraeul-ro, Jijeong-myeon, Wonju-si,
Gangwon-do, Korea

Certification ECM
Mark:



Product:
Model(s):

Thermal massage Bed
NM-5000P

Verification to:

Standard:
EN 60335-2-32:2003+A1:2008+A2:2015 used in
conjunction with EN 60335-1:2012+A11:2014,
EN 62233:2008, EN 55014-1:2017,
EN 55014-2:2015, EN 61000-3-2:2014,
EN 61000-3-3:2013

related to CE Directive(s):
2014/35/EU (Low Voltage)
2014/30/EU (Electromagnetic Compatibility)

Remark: The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Compliance Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. The regulation can be found at www.entecerma.it. This Certificate of Compliance can be checked for validity at www.entecerma.it

This verification doesn't imply assessment of the production of the product(s).

Additional information, clarification about the **CE** Marking:



We attest that a TCF for the **CE** Marking process is in place. Whereas the Manufacturer is Responsible to start the **CE Marking Certification Procedure** and to perform all the necessary activities, as required by the Directive before placing the **CE** Mark on the product(s).

Date of issue 23 January 2019

Expiry date 22 January 2024

Chief Manager
Marco Morino

Deputy Manager
Amanda Payne

Ente Certificazione Macchine Srl

Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it